

Government of Nepal
Ministry of Home Affairs
Armed Police Force, Nepal
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal



Amendment Notice

Date: 2081/08/28

1. Reference to our notice (IFB No: APFH/NCB/G/03/2081/082) Dated 2081/08/09 published on Prabhab Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

Changes made in Physiotherapy Relevant Equipments (Contract ID. No.APFH/NCB/G/03/2081/082/002)		
Combination Therapy Unit (Portable)		
S.No	Current Specification	Amended Specification
4.17	It should have at least 6" color touch screen.	It should have at least 5" color touch screen.

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/03/2081/082 shall remain the same.

Government of Nepal
Ministry of Home Affairs
Armed Police Force, Nepal
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal



Amendment Notice

Date: 2081/08/28

1. Reference to our notice (IFB No: APFH/NCB/G/03/2081/082) Dated 2081/08/09 published on Prabhab Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

Changes made in Neurosurgery Relevant Equipments (Contract ID. No.APFH/NCB/G/03/2081/082/003)		
Mayfield Three pin Skul Clamp		
S.No	Current Specification	Amended Specification
2.2	The system should have European CE or US FDA approved certificate.	The system should have European CE or FDA approved certificate.

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/03/2081/082 shall remain the same.

Government of Nepal
Ministry of Home Affairs
Armed Police Force, Nepal
Nepal A.P.F Hospital, Procurement Section
Balambu, Kathmandu, Nepal



Amendment Notice

Date: 2081/08/28

1. Reference to our notice (IFB No: APFH/NCB/G/03/2081/082) Dated 2081/08/09 published on Prabhab Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

Changes made in General Surgery Relevant Equipments (Contract ID. No. APFH/NCB/G/03/2081/082/004)		
Irrigator pump (for Urology, ENT, Laproscopic, Arthroscopic & other minimal invasive surgery)		
S.No	Current Specification	Amended Specification
1.13	Flow = Max 2200 mL/min	Flow =1200ml/min or more

Diode laser for Proctology/Varicose Vein		
S.No	Current Specification	Amended Specification
4.6	Fiber connection should be SMA905 international standard	Fiber connection should be available compatible with device.

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/03/2081/082 shall remain the same.

Government of Nepal
Ministry of Home Affairs
Armed Police Force, Nepal
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal



Amendment Notice

Date: 2081/08/28

1. Reference to our notice (IFB No: APFH/NCB/G/03/2081/082) Dated 2081/08/09 published on Prabhab Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

Changes made in Brain Scanner (Contract ID. No. APFH/NCB/G/03/2081/082/005)		
S.No	Current Specification	Amended Specification
7.1	ISO 13485 and EU or USFDA certificate must be submitted	ISO 13485 and EU or ISO 13485 or EU or USFDA certificate must be submitted.

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/03/2081/082 shall remain the same.

Government of Nepal
Ministry of Home Affairs
Armed Police Force, Nepal
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal



Amendment Notice

Date: 2081/08/28

1. Reference to our notice (IFB No: APFH/NCB/G/03/2081/082) Dated 2081/08/09 published on Prabhab Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

Changes made in 4k Tower (Contract ID. No.APFH/NCB/G/03/2081/082/007)		
S.No	Current Specification	Amended Specification
1.10	It should have multiple source management which should allow simultaneous control, display and documentation of image data from two connected image sources	It should have multiple source management which should allow simultaneous control, display and documentation of image data from two connected image sources/ tablet
1.15	Video Outputs: DVI-D, DP, 12G-SDI etc	Video Outputs: DVI-D, DP, 12G-SDI/ 4*3G-SDI
2.3	Camera head should have 3chip CMOS video sensor technology.	Camera head should have 3 chip CMOS or 4chip CMOS video sensor technology
4.1	Should be supplied with two laser/led light source	Should be supplied with two laser/Led Light source one for normal LED and other one for ICG activation
4.5	It must have a touch screen user interface	It must have a touch screen or ipad/tablet application user interface
Additional point		
The system should be fully ICG based camera processor and camera head should be ICG based.		
Telescope:-		
Diameter of approx 4mm and length of approx 180mm.		
Angle of view must be variable from 15 degree to 90 degree.		
Must be autoclavable with adjustment knob for changing the direction of view.		

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/03/2081/082 shall remain the same.