



Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal

Amendment Notice

Date: 2079/01/12

Reference to our notice (IFB No: APFH/NCB/G/04/2078/079) dated 2078/12/22 published on Rajdhani Daily Newspaper. This is to inform all concerned parties that some points of technical specification in bid document has been amended.

For detail amended notice please visit www.bolpatra.gov.np/egp or www.apf.gov.np. or visit Nepal APF Hospital, Procurement Section, Balambu Kathmandu or contact Tel 01-4315224.

All other terms and conditions of the IFB No. APFH/NCB/G/03/2078/079 shall remain the same.

Government of Nepal
Ministry of Home Affairs
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal
Amendment Notice
Date: 2079/01/12



1. Reference to our notice (IFB No: APFH/NCB/G/04/2078/079) Dated 2078/12/22 published on Rajdhani Daily Newspaper. This is to inform all concerned parties that following points of Technical Specification in bid document has been amended.

Changes made Cardiology Equipments (TMT & Holter ECG Machine) (Contract ID. No. APFH/NCB/G/04/001/2078/079)		
TMT Machine		
Tender point no.	Technical specifications as published in tender	Amended specifications
4.6	System should preferably have options to use both wire and wireless (wifi based) stress test	System should mandatorily have wireless stress test system and should be compatible for wired stress test for future use
4.7	It should have real time ECG transmission via WIFI	It should have real time ECG transmission via wireless technology (Wifi or Bluetooth)
4.9	Full set of TMT i.e. software, accumulators, treadmill, etc. should be of same company	Full set of TMT i.e. software, accumulators, treadmill, automatic BP recorder should be of same company
4.24 c	Have elevation range 0-40 degree (please specify)	Have elevation range 0-25 degree (please specify)
7.2	European CE (93/42 EEC directives) and USFDA approved product certificate. Relevant certificate should be attached for verification	European CE (93/42 EEC directives) AND/OR USFDA approved product certificate. Relevant certificate should be attached for verification
Holter ECG Machine		
Tender point no.	Technical specifications as published in tender	Amended specifications
2.1	Holter monitor with 3, 7 and 12 channel digital recorder with removable SD card as well as USB transfer and Bluetooth	Holter monitor with 3 and 12 channel digital recorder (derived 12 channel digital recorder will be accepted) with removable SD card as well as USB transfer and Bluetooth.
4.2	3, 7 and 12 channel arrhythmia analysis	3 and 12 channel arrhythmia analysis, derived 12 channel arrhythmia analysis will be accepted.
4.11	It should be water resistance IPX4 or better	It should be water resistance IP22 or better. Holter machine should have water protection facility. The higher the protection, the better and will be prioritized
4.16	It should have function to take a snap shot of a trend graph and add it to report	It should have feature of adding event picture to the report
4.17	Holter point no. 4.17 patient protection against electric shock hazard from USB cable > 3000 V. Holter system operates in not more than 4 volt with DC line. So this point should be omitted	This point has been removed.
4.23	Digital recorder must have a minimum of 512 samples/sec/channel for pacemaker spike detection and at least 128 samples/sec/channel for recording and storage of 1000/sec/channel for VLP	Digital recorder must have 512 samples/sec/channel or more for pacemaker spike detection and at least 128 samples/sec/channel or more for recording and storage of 1000/sec/channel or more for VLP.

All other terms and conditions of above mentioned **Contract ID. No. APFH/NCB/G/04/001/2078/079** shall remain the same.